

ASP Name/ASP Stamp: <b>JOSH TECH SERVICE PVT LTD.</b> CHENNAI. KANNAN. S-J 99000 12467.	<b>Report No.: LSI/2023/ 002681</b> <b>Service Toll-Free No : 1800 209 701</b> <b>Service Email ID:</b> <b>ininfo@thermofisher.com</b>
--	---

**Authorized Service Provider for ThermoFisher Scientific (LSI Division)**

**SERVICE / INSTALLATION REPORT**

CUSTOMER NAME	INDIAN IMMUNOLOGICALS LTD		VISIT DATE	20/07/2023	
ADDRESS	Kozipanni, Pudumudi, Ooty		ISM CASE NO.	327 749	
			MODEL NO.	FDE4004 OLV.	
CITY	Ooty	PIN CODE	643 007.	SERIAL NO.	11257745012205 02.
STATE	TN.				
CONTACT PERSON NAME/ DEPT. / DESIG.	T. SHANMUGA SUNDARAM QA DEPARTMENT.		INSTRUMENT STATUS	<input checked="" type="checkbox"/> WARRANTY <input type="checkbox"/> AMC <input type="checkbox"/> CMC <input type="checkbox"/> PAID <input type="checkbox"/> SALES SUPPORT	
TELEPHONE / MOBILE:	9942942102.		CALL REASON	<input type="checkbox"/> BD <input type="checkbox"/> PM1 <input type="checkbox"/> PM2 <input checked="" type="checkbox"/> INSTALLATION <input type="checkbox"/> CALIBRATION <input type="checkbox"/> VALIDATION	
EMAIL:	Shanmuga@indimmune.com.				

PROBLEM REPORTED:	Installation of -40°C Freezer.
-------------------	--------------------------------

SERVICE RENDERED: \* Unpacked and Inspected the Equipment. \* Found No Physical Damages. \* Inspected Incoming Voltage and Matching with Specifications. \* Connected with UPS Power Supply. \* Inspected Instrument Parameters. \* All Parameters Working Good. \* No Issues Found. \* Set Temperature Reached.

CUSTOMER REMARKS (IF ANY)	Equipment installation completed. QA, OQ and PQ yet to be performed.
---------------------------	--

ACTION REQUIRED	No
-----------------	----

JOB COMPLETED	<input checked="" type="checkbox"/> YES <input type="checkbox"/> NO
---------------	---

VISIT DETAILS			
DATE	FROM	TO	HOURS WORKED
		NA	

	KANNAN. S-J 
CUSTOMER FULL NAME / SIGNATURE STAMP	Engineer's NAME / Signature

THERMOFISHER SCIENTIFIC INVOICE NUMBER	
Notes (if any):	