

SERVICE REPORT

No. : _____

Date : 09/05/2019

To, Indian Immunologicals limited,
Grachibowli

Please acknowledge completion of the following services.

Your P.O NO : _____

Dated : _____

Sl.No.	DESCRIPTION OF SERVICE	UOM	REMARKS	CHARGE BASIS/FOC
1	Commissioning and calibration of O ₂ gas detector make Honeywell	01	Service Completed Status: <input checked="" type="checkbox"/> OK/Fail	Charge basis
			Service Completed Status: OK/Fail	
			Service Completed Status: OK/Fail	

For MMA CALABS TECH LTD

Service Witnessed By:

Name : S. Gopi

Name : _____

Designation : Service engineer

Designation : _____


Ph. No. : 9346004047/

Dept : _____

Ph. No. : _____

Email Id : calibration@mmacalabs.com

Email Id : _____

Signature : 

Authorized Signature: _____



ISO 9001:2008 Certified Company

Annexure to Service Report

To: Indian Immunologicals Limited
Gachibowli

No: _____

Date: 09/05/2019

Your PO No: _____

Dated: _____

Sl.No.	Detector Type/ Model	Sensor No	Transmitter No.	Span	L Alarm	H Alarm	Location/Tag	Condition
1	D2/XLD	K0154819111436	K0253419111436	20.9%	19.5%	23.5%	RD/D-09/Repository	OK

For MMA CALABS TECH LTD

Service Witnessed By:

Name: S. Gopi

Designation: Service engineer

Ph. No. : 9346004047/

Email Id : calibration@mmcalabs.com

Name : Satish Kumar Majumdar

Designation : Cr. Manager.

Dept : R&D

Ph. No. : 9700642870.

Email Id : m.sathishkumar@radimmune.com.

Signature : [Signature]
9/5/19

Authorized Signature: [Signature]
09/05/2019.