

SUN TEKNOLOZY / SUN LABORATORY SERVICES / SUN SCIENTIFIC CORPN

14/1 Lakshmi Dwaraka Villa, Thalayari Street, West Mambalam, Chennai-600 033, Ph.:044-24740143, Fax: 044-42024721
 Email: admin@sunteknoloji.com, sunteknoloji@gmail.com

**SERVICE REPORT
 CUSTOMER DETAILS**

Date: 16,17,18,29,30-10-18

| | | | |
|--|----------------|-----------------|--------|
| Customer Name & Address | Contact Person | MR. Venkatesan | |
| Human Biologicals limited, Cooty - 643007 | Designation | Manager | |
| | Dept.: | Quality control | |
| | Tel. No.: | O.: | Extn.: |
| | Mobile.: | | |
| | E-Mail.: | | |

INSTRUMENTS DETAILS

| | |
|------------------|-------------------|
| Instrument Name | SAS Air Sampler |
| Instrument Model | SAS Super ISO 100 |
| Sr. No | 16-03-3254 |

SERVICE DETAILS

Installation
 warranty
 AMC
 Services
 Calibration
 Chargeable/Free. if chargeable Amount Rs. _____


Problem Reported:
 New Installation of Microbial Air
 Sampler ISO 100

Action Taken:
 Supplied the SAS Super ISO 100, completed
 the installation of room air sampler. IQ, OQ and PQ
 document also completed. Given demo and installation of
 ISO manager software completed. Demo training and operation
 instrument working good condition.

SPARE PARTS REPLACED / RECOMMENDED

| Sr. No. | Item Code | Item | Qty. | Replaced | Recommended |
|---------|-----------|------------|------|------------|-------------|
| | | NA | | NA | |
| | | A-Santhosh | | A-Santhosh | |

Customer's Remark:
 Installation completed

| | |
|---|---|
| A-Santhosh 30.10.18 A-Santhosh Engineer's Signature | Machine Working Satisfactory  30.10.18 Customer's Seal & Signature |
|---|---|