


SERVICE REPORT

 CARE BIOSYSTEMS INDIA PVT. LTD. 8 & 8A, Ground Floor, Swaraj Daffodils, Plot No.96, Sector 8A, Airoli, Navi Mumbai - 400 708. Maharashtra State, India, Tel: +91-22-65569660/61. Email : service.mumbai@carebio.com		Date : 22/11/2018	JC#	719
		Engineer's name <i>Anthony Reddy + Mahipal Reddy</i>		
P.O. No. : 4511002602		Date of Request		Date of Visit 21,22/11/18
Customer's Name & Address <i>Mr. John Wesley .c</i>		Type of Services		
<i>Humman Biologicals Institute</i>		1 Installation <input checked="" type="checkbox"/>	6. Paid Service <input type="checkbox"/>	
<i>Kozhupuram</i>		2 Breakdown <input type="checkbox"/>	7. Calibration <input type="checkbox"/>	
<i>Pudumund Post.</i>		3 AMC <input type="checkbox"/>	8. IQ-OQ-PQ / Mapping <input checked="" type="checkbox"/>	
<i>Udhagamandalam. T.N. India</i>		4. Warranty <input type="checkbox"/>	9. Application <input type="checkbox"/>	
Tel. No. : Fax No. :		5. Inspection <input type="checkbox"/>	10. Goodwill <input type="checkbox"/>	
E-mail :		Estimated Amount Rs.		Add GST as applicable
Instrument	Model No.	Sr. No.		
<i>Haier Pharmaceutical Refrigerator</i>	<i>H4C-390</i>	<i>BEO6MKE17000EJ97000</i>		
<i>Carebio datalogger with Software</i>	<i>DL-35W DAS-V16.0</i>	<i>CB506101802</i>		
Request for / Fault : <i>Installation of Haier Refrigerator</i>				
Supply Voltage : <i>229 VAC</i>	Room Temp. : <i>25°C</i>	AC. : Yes <input type="checkbox"/>	No. <input checked="" type="checkbox"/>	
Stabilized supply : Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	Location : Room No. <i>CC</i>		Floor	
Instrument Status : Clean <input checked="" type="checkbox"/>	Average <input type="checkbox"/>	Not Good <input type="checkbox"/>	Needs Attn. <input type="checkbox"/>	
Set Temp. =	Actual Temp. =	Set RPM =	Actual RPM =	
Set =	Actual =	Set =	Actual =	
Work done : <i>Above the Refrigerator unpacked and shifted to the lab and installed. Performed TACOPO activity and along with installed Datalogger and Software found the satisfactory.</i>				
<i>Submitted the Document & Demonstration. Completed.</i>				
Spare Parts Used / Required : <i>—</i>				
Notes / Conclusion : <i>—</i>				
Job Completed : Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>				
Customer's Remarks : <i>IQ, OQ and PQ completed. System and Software performance will be observed for few days and performance will be updated to the vendors</i>				
Engineer's Signature <i>Anthony Reddy</i> 22/11/2018		Customer's Name & Signature : <i>A. VIJAY JOHNSON</i> 22-11-18		