



Towa Optics (India) Pvt. Ltd.,

C.4. 4th Floor, Gajel Apartments. 152. Greams Road, Chennai 600 006.

Ph: +91 44 28292010, 28291089. Email: towamds@towaoptics.com

INSTALLATION / SERVICE REPORT

Ref: TW.....

11249

Date

18/11/2021

Installation Service Visit ASC Visit 1 ASC Visit 2 Courtesy Visit PM Visit

System Under Warranty Yes No

INSTITUTION NAME & ADDRESS..... Human Biologicals Institute,

..... Kozhipannai, Pudumund Post, Coim - 643007

DEPT..... of QC - cell culture USER NAME..... Mr. Sharmugasundaram QA.

Phone:..... Date/s of Visit.....

EQUIPMENT..... Nikon Inverted M/S

MODEL..... TS-2 FL SI. No..... TS2-FL - 245886

ATTACHMENT(s)..... DS-Fi-2 / M.S-D.

ENGINEER(s)..... R. Velmurugan

INSTALLATION(s)

Received from..... Vide Invoice No..... Dt.....

ACTION TAKEN..... ET 510 Lp Bomies felter Replaced (New), New EPI issue

..... Solved, then over all equipment, optical-lens Fully cleaning

..... work has been done.

SPARES REPLACED.....

EQUIPMENT WORKING..... in good condition.

REMARKS..... Fluorescent Microscope working fine. We can distinguish Counter (Evans Blue) S-

..... as well as FITC stain clearly.

18/11/2021
ENGINEER'S SIGNATURE

18.11.21.
CUSTOMER'S SIGNATURE with SEAL