



# TRIDENT EQUIPMENTS PVT. LTD.

A - 105,, Kailas Industrial Complex, Hiranandani Godrej Link Road, Vikhroli (W),  
Mumbai - 79. Tel.: 2518 1704 / 05 / 06 Telefax : 2518 5463 E-mail : info@tridentequipments.com

## Service Report

No. 61979

Name of Customer: <u>Human Biologicals Institute.</u>	Product: <u>pH-Meter.</u>
Address: <u>Oray, Tamilnadu.</u>	Model(s): <u>Oxon Starc-291</u>
Contact Person/Dept: <u>D. VENKATESAN</u>	Serial No.(s): <u>X-73056</u>
PO No.: <u>4511004019</u>	Installation <input checked="" type="checkbox"/> Warranty <input type="checkbox"/> Chargeable <input type="checkbox"/> Courtesy <input type="checkbox"/>
Telephone: <u>80129 55566</u>	Repaired at - Customer Site <input checked="" type="checkbox"/> Service Centre <input type="checkbox"/>
Email:	AMC : Preventive Maintenance Visit No. <u>—</u>
	Breakdown Visit No. <u>—</u>
	Date: <u>13.12.2022.</u>

### Nature of Problem :

Installation of pH-meter.

Electrode → 8157 BNUMD.  
AX1-2012.

Action Taken :

- 1) Meter has been calibrated and verified using simulator and buffer. Calibration is performed. Results found within acceptable limits.
- 2) Temperature calibration performed and found satisfactory.
- 3) IQOQ documentation is done and filled as per PO.
- 4) Meter and electrode is ready to use for analysis.

### Engineer's Comment :

Installation and IQOQ of pH-meter is completed.

### Customer's Comment :

Installation completed successfully and equipment working satisfactorily. Related documentation and training completed.

Service Engineer's Name & Signature : Jayashree  
13.12.22.

Customer's Name, Signature & Stamp : [Signature]  
13.12.22

Encl.

Mob: → 8249698697

# TRIDENT EQUIPMENTS PVT. LTD.



Consumable Detail's			Standard Details		
Due Date	Next Due		Lot No.	Due Date	Next Due
UV Lamp	<input type="text"/>	<input type="text"/>	Calibration	<input type="text"/>	<input type="text"/>
Pump Head	<input type="text"/>	<input type="text"/>	Verification	<input type="text"/>	<input type="text"/>
Resin Bed	<input type="text"/>	<input type="text"/>	SST	<input type="text"/>	<input type="text"/>
Pump Tubing	<input type="text"/>	<input type="text"/>			
Tubing Set	<input type="text"/>	<input type="text"/>			

Service Engineer Name & Signature

Customer Name, Signature & Stamp

*Meter has been calibrated and verified. Results found satisfactory. Calibration is performed.*

*Installation completed successfully and equipment working properly. Report documentation and training completed.*