

Eppendorf ThermoMixer C/F, ThermoStat C, MixMate

Installation report

Copyright ©2019 Eppendorf AG, Germany. All rights reserved, including graphics and images. No part of this publication may be reproduced without the prior permission of the copyright owner.

Eppendorf® and the Eppendorf Brand Design are registered trademarks of Eppendorf AG, Germany.

Registered trademarks and protected trademarks are not marked in all cases with ® or ™ in this manual.

1 Installation report

Customer and device information

Manufacturer	Eppendorf AG
Device name	Thermomixer C
Serial number of the device	5382KJ740991
Inventory number of the device	NA
Date of installation (year, month, day)	2021-12-10
Company name	Human Biological institute
Company address	Indian Immunologicals Limited, Kozhipannai, Pudumund PO,643007 Ooty, Tamil Nadu
Location of the device: Address	Indian Immunologicals Limited, Kozhipannai, Pudumund PO,643007 Ooty, Tamil Nadu
Location of the device: Building	QC
Location of the device: Floor number	1
Location of the device: Room number	QC
Location of the device: Lab number	QC

Document history

Version	Change	Date
Version 00	Document created.	2019-November-17

Installation and initial setup of the device

<input checked="" type="checkbox"/> The ambient conditions were checked.	
<input checked="" type="checkbox"/> The delivery was checked for completeness.	
<input checked="" type="checkbox"/> The packing was checked for damage.	
The device has been connected to the	<input checked="" type="checkbox"/> Mains/power line <input checked="" type="checkbox"/> Emergency shut-off (if applicable) <input type="checkbox"/> Data network (if applicable)

The checked off inspections were performed. Any deviations must be documented in the "Comment" section.

Installation report

Eppendorf ThermoMixer C/F, ThermoStat C, MixMate
English (EN)


Function test

The following functions were checked	<input checked="" type="checkbox"/> Speed setting (if applicable) <input checked="" type="checkbox"/> Temperature setting (if applicable) <input checked="" type="checkbox"/> Timer (if applicable) <input checked="" type="checkbox"/> SmartBlock detection (if applicable) <input checked="" type="checkbox"/> Display function <input checked="" type="checkbox"/> Function of the operating controls
--------------------------------------	---

The checked off inspections were performed. Any deviations must be documented in the "Comment" section.






Comments

Installation and demonstration done working fine
Warranty 1 year from date of installation.
The Thermomixer equipment No QC/TMIX/203 was allocated on 20.12.21. The Thermomixer qualification and documentation of IQ and OQ was completed. The performance of the equipment was found satisfactory. Hence, the equipment shall be released for routine use.


20.12.21


Instructed persons


The following persons have been instructed in operating the device:

Surname and name	Function	Signature
T. Shanmugasundaram	QC	
P. Ambrose Antony Raj	QC	
M. Gokeshath	QC	
P. Senthil Kumar	QA	
M. Arun	Engineer	

Signature of the service technician

The device has been setup and commissioned in accordance with the manufacturer's specifications and can be operated according to its intended use.

Surname and name	Date	Signature
Kumar, Sudhi Suresh	10-12-2021	


10.12.21

Thank you!

www.eppendorf.com