

# MEDISPEC (I) LTD

Offering Product for future of Science ...

KH03/Elec/004.

## SERVICE REPORT

<input type="checkbox"/> INST	<input type="checkbox"/> PMS	<input type="checkbox"/> CL	Date:
			12-OCT-2021

AHMEDABAD  
  BANGALORE  
  CHANDIGARH  
  CHENNAI  
  DELHI  
  HYDRABAD  
  KOLKATA  
  LUCKNOW  
  MUMBAI  
  PUNE

Institution:	Indian Immunologicals Limited Karkapatla	Attn:	MV. Venkatarao Vemula
City:	Telangana	Pincode:	502281
Order Ref.:	4511003642	Tel.:	+91 7702014376
Model:	5075 8/12	Serial No.:	21080918
Base code:	1550200	Version:	1.05
Assaycode:		Assaycode:	
Version:		Version:	
PO Date:	29-June-2021	Email:	v.venkatarao@indimmune.com

Principals	<input checked="" type="checkbox"/> BioTak	<input type="checkbox"/> Syngene
	<input type="checkbox"/> NuAire	<input type="checkbox"/> Symbiosis
	<input type="checkbox"/> GeSIM	<input type="checkbox"/> WTA
	<input type="checkbox"/> Biospherix	<input type="checkbox"/>

Service Complaint:	Qualification of BioTek 5075 microplate washer.	Service Type	<input type="checkbox"/> Inspection <input checked="" type="checkbox"/> Installation <input type="checkbox"/> U / Warranty Service <input type="checkbox"/> U / AMC Service <input type="checkbox"/> Preventive Maintenance <input type="checkbox"/> Paid Service / Calibration <input type="checkbox"/> Add-on / Complimentary
--------------------	---	--------------	---

Observation:	-	Utilities Check List	<input checked="" type="checkbox"/> Electric Supply : 230VAC <input checked="" type="checkbox"/> Ground / Earthing <input checked="" type="checkbox"/> Voltage Stabiliser / UPS <input checked="" type="checkbox"/> Air Conditioning <input checked="" type="checkbox"/> Dust Free Environment <input type="checkbox"/> Computer <input type="checkbox"/> Printer <input type="checkbox"/> Interface : Serial / USB <input type="checkbox"/> COM Port : <input type="checkbox"/> Gas Supply <input type="checkbox"/> Distilled Water
--------------	---	----------------------	--

Service Details:	Completed 20-02-2021 All test results are PASS. Documents handover.
------------------	---

Details of Test Equipments / Test Jigs / QC Test Solutions / Materials / Parts used :					
No.	Part No.	Serial No.	Batch No.	Name / Description	Valid upto Date
1	7773001	NA	NA	Blue Test Dye	NA

Test Result:	Completed 20-02-2021 of microplate washer.
--------------	--

Note / Remarks:	
-----------------	--

For MEDISPEC (I) LTD.:	For Customer's Approval / Acceptance:
Engineer's Name:	Signature & Seal:
Dr. Satish B 12.10.2021 	Venkat 12-10-2021