

ULTRA INSTRUMENTS

Plot No.92, Amariyothi Colony, New Bowenpally,
Secunderabad - 500 011.
Tel: +91 - 40 - 27951055, 27951066
Email: us@ultragroup.in

ULTRA

www.ultragroup.in
GSTIN : 36AACFU7794F1ZQ
SERVICE TAX NO.: AACFU7794FSD001

Test Certificate

Ref: UIS/CAL/2226/TEMP

Date:28.09.2022

Customer Name:
**INDIAN IMMUNOLOGICALS LIMITED, KARKAPATLA
TELANGANA**

Test Equipment: Dry Block Heater

Make: IKA

Model: DBH2

Test Instrument Range: Ambient +5 to 500 °C

Serial No: 06.254091

Calibration Date:28.09.2022 & Due Date:27.09.2023

Accuracy: ±1 °C

Test Details: Test for the Temperature

Test Location: On- Site

External

TEST CONDITIONS

RECALIBRATION DEVICE DETAILS

Ambient temp: 24.5 °C

Calibration device Serial No: 222304555

Relative humidity: 51% RH

Device: Scan log 16Channel, Range: -199.9 to +600 °C

Make: PPI

Re-Calibrated by: PPI

Calibration validity: Valid from- 23/05/2022

Calibration validity: Valid to- 22/05/2023

Sr.No	Set Values (Temp °C)	Observed Values (Temp °C)	Variance (Temp °C)	Acceptance levels
1.	35	35.1	0.1	±1 °C
2.	55	55.3	0.3	±1 °C
3.	75	75.6	0.6	±1 °C
4.	95	94.5	0.5	±1 °C

TEST RESULT: ACCEPTED

REJECTED

Test Performed By: G Nithin

Date:28.09.2022


Authorized Signature.