

Installation/Service Report

M/S. : Indian immunologicals limited	Date : 08/08/2022	Report No: SAS/Inv 23
Address: Sr. NO 281 - 284 & 321 Karkandla Vill) Markok (M) Siddipet (D) TS 502281	Type of Visit / Job :-	
	<input checked="" type="checkbox"/> Installation & Demonstration <input type="checkbox"/> Repair / Service <input type="checkbox"/> Inspection <input type="checkbox"/> Calibration To Be Performed	
Tel. : 77020 14336		
Email ID :		
End User : Mx.V.Venkaateshwarara Rao		
Department : microbiology		
Instrument : Microbial Air sampler	Serial No. : 22-D-17915, 21-D-17624	
Model No. : SAS Super 100	20-D-16241, 20-D-16261	
Symptoms : New air sampler Installation.	20-D-16264	
Action Taken : Installation completed successfully. performe Qualifications as per the Qualifications IQ, OQ and PQ submitted all qualifications documents and other relevant documents to customer.		
Engineer's Remarks : NA		
Customer's Remarks : NA		

Engineer's Name: A.Siddi Ramiya

Signature

ASR 08/08/2022

End-User's Name: V.Venkaateshwarara Rao

Signature

Venka 08.08.2022