


SERVICE REPORT

CUSTOMER NAME : Indian Immunologicals S.NO. 32246 DATE 28/01/2021
 ADDRESS : Cochinbadi CASE NO. 162840
R&D Dept MODEL NO. Non-drop one
 CITY Hyderabad STATE Telangana SERIAL NO. AZY2021022
 PIN CODE : OTHER DETAILS OF THE UNIT
 CONTACT PERSON : Dr. Shantha Kumar P INSTRUMENT STATUS: Warranty Contract Billable
 TELEPHONE / MOBILE : 99515826202 CALL REASON: Breakdown Installation Preventive Calibration Application
 E-MAIL : Shanthakumar@indimmune.com Division: CMD LSI CAD EPM MSD SDG
 ALL DETAILS : Installation & Calibration & Demonstration

DETAILS OF SERVICES RENDERED : * Checked Consignment and found ok as per packing list * checked power supply and found ok.
 * Turned on Instrument and observed initializing without any error. * performed factory calibration and found ok with PV-1 solution. * Customer Demo given. * checked customer samples. * Instrument working satisfactorily.

NO.	PART#	SPARES DESCRIPTION	QTY. REPLACED	REPLACED FROM (✓)	
				CUST. STOCK	THERMO STOCK
<p>VIA <u>Vinnu</u> <u>28/01/2021</u></p>					

REMARKS (IF ANY) :
 JOB COMPLETED : YES NO

CALL TIME				CUSTOMER'S FULL NAME/SIGNATURE STAMP	Engineer's Name / Signature
Date	From	To	Hrs. Worked		
<u>28/01/2021</u>	<u>9:30</u>	<u>12:30</u>	<u>3:00</u>	 P. [Signature]	<u>Vinnu</u> <u>28/01/2021</u>

ACTION REQUIRED: YES NO