

SAKSHAM TECHNOLOGIES PVT. LTD.

SERVICE REPORT : SER/02/R1



108-110, Agarwal B2B Commercial Complex,
Dhruv Park, Kanchpada, Off. Link Road,
Malad (West), Mumbai - 400 064.
Tel. : (022) 2844 3636 / 2889 3535 / 2880 3636
Fax : (022) 2889 3636 / 2880 3738
Email : info@saksham.co.in / service@saksham.co.in
Website : www.saksham.co.in

No : **5802**

Customer's Name : Human Biological Institute
Address : Unit off IIL ooty.

Attended By : Ragupathi E

Phone No. :
Email ID :
Department : QC Microbiology (CTL)

Date : 06/11/2020

CRM - CL No. : 4632

Nature of Visit : Installation / Warranty / Paid Service / Service Contract / Demo / Validation / Inspection / F. S. / C. V.

Equipment : <u>LSS</u> <u>Safety box</u>	Model No. : <u>AC 600 CM</u>	Serial No. : <u>1348</u>
---	------------------------------	--------------------------

Service Rendered :

→ Above mentioned equipment flexible hose
4" & 4.5" provided and ducting done from
Cabinet to Exhaust and Exhaust blower to
outlet
→ Installation done successfully.

Our Remarks :

Ducting pipe connected to Installation done

Spares Parts Used :

4" to 4.5" flexible hose

Recommended :

Customer Remarks : warranty certificate yet to be provided

How Do You Rate Our Service :- 1) Not Good 2) Satisfactory 3) Good 4) Very Good 5) Excellent

Your Service Engineer has visited and carried out the work to my satisfaction.

I/We agree to pay the charges if any.

Customer's Name : D. VENKATESAN

Signature : [Signature] 06-11-20

Engineer's Signature : [Signature] 06/11/2020

Designation : MANAGER-OC

Mobile No. : 8012755566

Internal Use :

Job Completed Yes No

In Time : 10 AM Out Time : 5 PM

Courtesy Calls :-