

SERVICE REPORT

Our Ref. : <u>LCG/AS/LEAD/2020</u>	Engineer's Report No. <u>1662</u>	Principals <u>LCGCTrucal</u>
Customer <u>M/s. Padma Pharmaceutical Pvt. Ltd.</u>	Invoice Details : <u>—</u>	Activity
Address <u>Kankapatla (V).</u>	Instrument Details : <u>Analytical</u>	Installation : <input checked="" type="checkbox"/>
Scientist <u>Mr. Ashwini</u>	Model : <u>AS8220-X2</u>	Warranty : <input type="checkbox"/>
Phone _____ Fax _____ Telex _____	S.No. : <u>641035</u>	Contract : <input type="checkbox"/>
		Out of Warranty : <input type="checkbox"/>
		Billable : <input type="checkbox"/>

Work Details : Installation

- ① Installed the AS 8220-X2 Analytical Balance.
- ② Calibrated the Balance with standard weights.
- ③ Explained all functions to Concern Persons.
- ④ I/O O/P Documentation handed over to Concern Person.
- ⑤ Balance is working in good condition.

PARTS REPLACED

Part Number	Description	Quantity	Source	Amount
/	/	/	/	/

The service requested by us has been handled by Ramesh
Who visited us on 05/10/2020

He has carried out the above work to our satisfaction.

Customer Report :

Place : Kankapatla Engineer Signature : [Signature] Customer Signature : [Signature]
(with Seal) 05-10-2020

Date : 05/10/2020