

SERVICE REPORT

Our Ref.: LCGC-RADWAG/203/12-13			Engineers's Report No. 203	Principals
Customer <u>Indich Immunological Ltd</u>			Invoice Details :	LCGC-RADWAG
Address <u>Rachapuram Gachibowli</u>			Instrument Details : <u>Analytical</u>	Activity
Scientist <u>Mr. D. Phanikumar</u>			Model : <u>As 220/X</u>	Installation : <input checked="" type="checkbox"/>
Phone	Fax	Telex	S.No.: <u>382266/13</u>	Warranty : <input type="checkbox"/>
<u>9010722122</u>			(Analytical Bio Chemistry Lab R&D)	
			Contract : <input type="checkbox"/>	
			Out of Warranty : <input type="checkbox"/>	
			Billable : <input type="checkbox"/>	

Work Details :

- Below mentioned the balance details
- Checked the balance, it is found ok
- Installed the balance with above s.no
- Done the Internal & External Calibration
- Executed the IQ/OQ/PQ documentation and hand over it to concerned person
- Explained the all features of the balance to the end user
- Finally the balance is working in good condition

PARTS REPLACED

Part Number	Description	Quantity	Source	Amount
		1		

The service requested by us has been handled by Ch. Acharya
Who visited us on 04/03/13

He has carried out the above work to our satisfaction.

Customer Report :

Place : Gachibowli

Engineer Signature : [Signature]

Customer Signature : [Signature]
(With Seal) 4/03/13

Date : 04/03/13