

Plot No. 57, Sheshachala Colony,
Opp. Sarojini Naidu Memorial High
School, West Marredpally,
Secunderabad - 500 026, India.

SERVICE REPORT



Phone : (040) 4241 4328
Fax : (040) 4241 4327
E-mail : hyd@toshivin.com
Web : www.toshivin.com

Customer : Indian Immunological Limited	Product : Shimadzu UV-vis spectrophotometer	SRQ No.:
Contact : Mr. Srinivasa	Model No. : UV-1800 Sr. No. A11635581115	Call Type : ID/CR/WR
Dept. : Production	Arrival Date : 02/03/12 Dep. Date : 02/03/12	FR/APP/S/C
Address & : Gaekli Bouda	Time In : : Time Out : : Total Days :	
Tel. No. : 9905352225	Service Contract Visit No.	Revisit Required : <input checked="" type="checkbox"/> Y / <input type="checkbox"/> N
Fax No. : :	Product working satisfactorily <input checked="" type="checkbox"/>	Date of Revisit Committed :
E-mail : :	Send to Service Div. <input type="checkbox"/>	Beyond Repairs <input type="checkbox"/>
SYMPTOM : Installation	<p>REPAIR ACTION</p> <p>performed - the qualification (58000), checked and found the result satisfactory, completed the documentation, handed over to customer, Given the demo to the users, informed working satisfactory.</p>	
Customer's Remarks : Spectrophotometer (B31UV1162) - 5000 Completed.	<p>ACTION REQUIRED BY :</p> <p><input type="checkbox"/> Customer <input type="checkbox"/> Engineer</p>	
Consumed	Components To be Indented	Source
Service Engineer's Name : P. Ramel		Work Accepted by : P. Srinivasa
Signature : [Signature]		Signature : [Signature]

Plot No. 57, Sheshachala Colony,
Opp. Sarojini Naidu Memorial High
School, West Marredpally,
Secunderabad - 500 026, India.

SERVICE REPORT



Phone : (040) 4241 4328
Fax : (040) 4241 4327
E-mail : hyd@toshivin.com
Web : www.toshivin.com

Customer : Indian Immunological Limited	Product : Shimadzu UV-VIS Spectrophotometer	SRQ No.:
Contact : Mr. V. Srinivas	Model No. : UV-1800	Call Type : JD/CR/WR
Dept. : HBS Production	Arrival Date : 09/03/12	Dep. Date : 09/03/12
Address & : Godikonda	Time In : -	Time Out : -
Tel. No. :	Service Contract Visit No.	Revisit Required : <input checked="" type="checkbox"/> Y/N
Fax No. :	Product working satisfactorily <input type="checkbox"/>	Send to Service Div. <input type="checkbox"/>
E-mail :	REPAIR ACTION	Beyond Repairs <input type="checkbox"/>
SYMPTOM : Installation	Completed the installation with PC and printer, checked and found it working satisfactorily, O&O operations & Qualification is pending.	
Customer's Remarks	ACTION REQUIRED BY :	
	<input type="checkbox"/> Customer	<input type="checkbox"/> Engineer
Consumed	Components To be Indented	Source
Service Engineer's Name : P. Ramesh		Work Accepted by : M. A. S. Prasad
Signature : [Signature]		Signature : [Signature]