

LCGC Trical and Services LLP

Plot No. 57, Road No. 5, ALEAP Industrial Estate,
Gajularamaram (V), Quthbullapur (M), Hyderabad - 500090, T.S. India
Tel : +91 7995065246, 9246154047
E-mail : service@lccgindia.com
Website : www.lccgweighing.com

SR - 1870

SERVICE REPORT

Our Ref. : <u>LCGC/Radwag/1870/2025</u>	Engineer's Report No. <u>1870</u>	Principal <u>Radwag</u>
Customer <u>M/s. Indian Zirconological Ltd</u>	Invoice Details: <u>—</u>	Activity
Address <u>Karkapadla</u> <u>Hyderabad</u>	Instrument Details : <u>Precision</u>	Installation : <input checked="" type="checkbox"/>
Scientist : _____	Model : <u>PS 6100-02-M</u>	Warranty: <input checked="" type="checkbox"/>
Phone <u>—</u> Fax <u>—</u> Telex <u>—</u>	S. No. : <u>820679</u>	Contract: <input checked="" type="checkbox"/>
		Out of Warranty : <input checked="" type="checkbox"/>
		Billable : <input checked="" type="checkbox"/>
	<u>Inst No:- KA02/WDL/006</u>	

Work Details :

Installation:-

- * Installed the Balance in proper way.
- * Done the Internal/External calibration.
- * Performed the 20, 0K and 100g qualification with respective procedure, the results are with in the limits.
- * Done the zero to end use.
- * Now Balance is working in good condition.

PARTS REPLACED

Part Number	Description	Quantity	Source	Amount
<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>

The Service requested by us has been handled by O. Narendra + Shiva
who visited us on 03/04/2025

He has carried out the above work to our satisfaction,

Customer Report :

Place : Karkapadla Engineer Signature : [Signature] Customer Signature : [Signature]
Date : 03/04/2025 (with Seal) 03/04/2025