

SR - 462

SERVICE REPORT

LCGC TRUCAL AND SERVICES LLP

Plot No :57, Road No:5, ALEAP Industrial Estate,
Gajularamaram (V), Quthbullapur (M), Hyderabad-500090. T.S. India
Tel.: +91 7995065246, 9246154047
E-mail : service@lcccindia.com
Website : www.weighing@lcccindia.com

Our Ref. : <u>LCGC Padureg / 462 / 2023</u>	Engineer's Report No. <u>462</u>	Principals
Customer <u>Indian Environmental & Safety Ltd</u>	Invoice Details : <u>-</u>	Activity <u>Repair</u>
Address <u>S.No. 281 To 284 Cross, Kankapalle</u>	Instrument Details : <u>Precision</u>	Installation : <input checked="" type="checkbox"/>
<u>Frankend (M), Siddipet</u>	Model : <u>P5360-A2</u>	Warranty : <input checked="" type="checkbox"/>
Scientist : _____	S.No. : <u>9519921</u>	Contract : <input checked="" type="checkbox"/>
Phone : _____ Fax : _____ Telex : _____		Out of Warranty : <input checked="" type="checkbox"/>
		Billable : <input checked="" type="checkbox"/>

Work Details :

Installation :-

Installed the balance in laboratory.

Performed the zero, OB and PQ qualification with respective procedure. The results are with in the limit.

Done the zero to end user.

Now balance is working in Good condition

PARTS REPLACED

Part Number	Description	Quantity	Source	Amount
 	 	 	 	

The service requested by us has been handled by D. Haripandya Saiju

Who visited us on 18/04/2023

He has carried out the above work to our satisfaction.

Customer Report : _____

Place : Kankapalle Engineer Signature : _____ Date : 18/04/2023

Customer Signature : _____ (with Seal) K. TRIF 18/04/2023