

## VICE REPORT

Report. No:

Organisation : Indian Immunologicals Limited. Chikowli - 500032, Hyderabad		Engineer : M. Shadw Shetty.	
Date : 12-01-23, 13-01-2023		PO.No: 4511004101	
Customer Name : P. Sulkeema.		Nature of Visit	
Designation : Manager.		Installation <input checked="" type="checkbox"/> Warranty <input checked="" type="checkbox"/>	
Department : Production		Paid Service <input type="checkbox"/> AMC <input type="checkbox"/>	
Email : P.Sulkeema@indimmune.com.		Courtesy <input type="checkbox"/> Others <input type="checkbox"/>	
Phone no : 9490065900.			
Equipment : CO Biosafety Cabinet	Model : AC2-USB-NS	Serial.no 2022-192349 Evid:- B5/BSC/248 2022-192348 Evid:- B5/BSC/249	
Complaint: Installation			
Work carried out: Above mentioned equipment inspection done, performed. Installation Qualification, Operational Qualification, Performance Qualification done, found all parameters within the limits. Reference documents & certificates handed over to customer. Equipment is ready to use for regular activity.			
Instruments used: — NA —		Recommendations: — NA —	
Dealer Remarks: Equipment is ready to use for regular activity			
Customer Remarks: CD - Service pattern to be given by NVAK; & DA document to be given			
Customer Signature P. Sulkeema 13.01.23		Engineer Signature M. Shadw Shetty 12.01.2023	

Completed: Yes  No