



Engineering Installation Report

Name of Institution: Indian Immunologicals Limited
 Door No: - Building Name: R&D Floor: Ground Floor
 Street/Road: Rakshapuram Colony/Sector/Nagar: opp Gachibowli Stadium
 Location: Gachibowli City: Hyderabad Pin Code: 500032 Phone: 9030465523
 Department (kindly ✓): Pathology Blood Bank Other (mention) R&D Lab

Person In Charge		Instrument Main Contact	
Name	<u>A. Praveen</u>	Name	<u>A. Praveen</u>
Phone	<u>9030465523</u>	Phone	<u>9030465523</u>
Designation	<u>Sr. Manager</u>	Designation	<u>Sr. Manager</u>
Email	<u>apraveen@indimmune.com</u>	Email	<u>apraveen@indimmune.com</u>

Instrument Serial No & Model Name: Cedar Bio S.NO - 36526
 Software Ver: 3.0.0.2130 Cobas Link Serial No: -
 Fortigate Make / Model: - Fortigate Serial No: -
 Installation Commencement Date: 28/03/2025 Date of Installation Completion: 28/03/2025
 Travel Hours 2 hour. Work Hours 4 hour. Work ORD No -

◆ **Transportation Damage & Discrepancy Report: {please update after inspection}**
 A. List of Missing items: NIL
 B. Notes on any damages: NIL
 C. General Remark: NIL

◆ **Installation checks & Summary: {please indicate status}**

A. Power Supply:	<u>220V</u>	D. Fluidic Adjustment & Setup:	<u>done</u>
B. Software Installation & Boot-up:	<u>done</u>	E. System Checks & Status Print Outs:	<u>done</u>
C. Mechanical Adjustments & Setup:	<u>done</u>	F. Other (If any, pls mention)	<u>done</u>

Customer Comments: _____

Signature of Engineer [Signature] Signature & Seal of Customer Praveen Date 28/3/2025

Name of Engineer G.C. Krishan Reddy Engineer's Comment Analyzer Working Fine.

Note: Instrument main contact from Lab shall be responsible to receive act and send acknowledgement related to all field notification from Roche. Customer shall be responsible to notify Roche at its customer care in case of any change for Instrument Main Contact to receive timely information.

"In case of Complaints/Inquiries please reach our Customer Support Centre (CSC, Customer Care) at 1800-123-7599 (Toll Free) / 022-66974755"