

SERVICE REPORT

CUSTOMER NAME : Human Biologics Institute S.NO. 19268 DATE 12-02-21
 ADDRESS : Indian Immunologicals Limited CASE NO. _____
 Rakshapuram, Kachibowli MODEL NO. SORVALL ST16R
 CITY Hyderabad STATE Telangana SERIAL NO. 42652472
 PIN CODE : 500032, Block-B2 OTHER DETAILS OF THE UNIT B2/CN/068
 CONTACT PERSON : Sri Krishna INSTRUMENT STATUS: Warranty Contract Billable
 TELEPHONE / MOBILE : 1714 / 1715 CALL REASON: Breakdown Installation Preventive Calibration Application
 E-MAIL : s.krishna@indimmune.com Division : CMD LSI CAD EPM MSD SDG

CALL DETAILS : Installation and IQ, OQ, PQ of Sorvall ST16R centrifuge



DETAILS OF SERVICES RENDERED :

- unpacked the consignment and verified that all items are received under good condition as per the good purchase order.
- Installed the instruments and all its accessories
- connected the instrument to the UPS power
- performed the IQ, OQ, PQ of the instrument
- All measurements are under acceptable criteria.
- handed over the documents & reports to the customer
- demonstrated the operation of instrument to the users.
- completed the installation & qualification, instrument is working fine

NO.	PART#	SPARES DESCRIPTION	QTY. REPLACED	REPLACED FROM (✓)	
				CUST. STOCK	THERMO STOCK
/					

REMARKS (IF ANY) : _____

JOB COMPLETED : YES NO

CALL TIME				CUSTOMER'S FULL NAME/SIGNATURE STAMP	Engineer's Name / Signature
Date	From	To	Hrs. Worked		
				 12-02-21	 12-02-21

ACTION REQUIRED: YES NO