

Hanna Equipments (India) Pvt. Ltd.

4/5/6, Aum Sai Building, Plot No. 23C, Sector 7, Kharghar, Navi Mumbai - 410210.
 Tel: +91-22-27746554 / 55 / 56, Fax: +91-22-27746557
 Email: tech@hanna-india.com, Website: www.hannainst.com/usa
 Service Tax Regn: AACCH2670QST001 VAT TIN: 27120729904



INSTALLATION/ DEMONSTRATION/ SERVICE REPORT

Company: <u>Indien immunologicals</u>	Instrument: <u>HI5321-02</u>	Report No.: <u>2078</u>
Contact Person: <u>Mr. Anjeyulu. L.H.</u>	Type: <u>conductivity</u>	Installation <input type="checkbox"/>
Department: <u>production</u>	Model No.: <u>HI5321-02</u>	Demonstration <input checked="" type="checkbox"/>
Address: <u>Gaechibowli</u>	Serial No.: <u>F0106659</u>	Warranty <input type="checkbox"/>
<u>Hyderabad</u>		Out of Warranty <input type="checkbox"/>
<u>Telangana.</u>		AMC <input type="checkbox"/>
Tel: _____		
E-mail: _____		

Date of visit: <u>22/02/2018</u>	Engineer: <u>J. Jay2.</u>	No. of Days/Hrs.: <u>1</u>
----------------------------------	---------------------------	----------------------------

Nature of complaint: Installation of HI5321-02

Action taken: calibration standards checked 84, 1413, 1288 successfully and WFI = 0.68 is observed. sample.

Remarks: calibration done successfully with standards, and samples also checked with accurate results.

Parts replaced/ required				Charges	
S. No.	Description	Quantity	Amount	Spares charges:	_____
				Service charges:	_____
				Tax:	_____
				Total:	_____

Customer remarks: Required calibration certification and instrument installation certification separately. other wise bill not process.

The
 Demonstration
 Installation
 Service
 Has been carried out and completed to our satisfaction

Customer sign with date: Sri 22.02.18

Name of the Engineer: J. Anjanulu
22/02/2018

Engineer sign with date: [Signature]