



MICROFLOW DEVICES INDIA PRIVATE LIMITED

Regd. Office : Plot No : 4 & 5 WIP Sidco Industrial Estate, Thirumudivakkam, Chennai - 600 044. Ph : 91-44-24781010 / 95000 82797

Correspondence Address : No. 52G, 1st Cross Street, Sabari Nagar, Porur, Chennai - 600 116.



INSTALLATION / SERVICE REPORT

No. :

Date : 05/05/17

Client : Indian Immunologicals limited

Location : Rashtapuram, Kankapatta,

Contact Person : Mr. Nataraj

Department : _____

Nature of work : Installation Modification AMC

Warranty General Service Others

Model	Sr.No	*Service rendered	Average velocity fpm	Pressure mm of water
MFD/FT/1500	2585/02/2017	Fume Hood Installation work is completed.	-	-
<p>Remarks</p> <ol style="list-style-type: none"> 1) Sink provided by acrylic sheet but URI maintain epoxy 2) IO, OQ, PQ & DQ not provided 3) Adjust not properly sealed. Edges 				
<p><u>[Signature]</u> 9500120047 Rajesh P / B. Gresh Babu.</p> <p>Customer Support Engineer</p>		<p>Client's Remarks</p> <p><u>[Signature]</u> 05-05-17</p> <p>Date Signature</p>		