



INSTALLATION REPORT

DELONIX HEALTHCARE PVT LTD FLAT NO:6 PLOT NO:203 RAO'S ENCLAVE 5 MARUTI NAGAR, MOULA-ALI HYDERABAD - 500 040.		DATE: 04.04.25		S.NO:	
		ENGINEER'S NAME: Nagendra Reddy K			
		MOBILE NO: 9848199139 & 9515629688			
CUSTOMER'S NAME			TYPE OF SERVICES		
M/s Human Biologicals Institute Raksha Puram, Gachibowli Hyderabad PO No- 4511005027			INSTALLATION BREAKDOWN AMC WARRANTY INSPECTION	Yes	PAID SERVICE CALIBRATION IQ/OQ/PQ APPLICATION GOODWILL
TEL. NO:	FAX. NO:	JOB COMPLETED	YES NO		
E MAIL:					
INSTRUMENT: Freeze Dryer		MODEL NO. FD-10-TP		S. NO./INST. ID 2411FD-27TP	
PROBLEM AS REPORTED BY THE CUSTOMER: SERVICE DONE: Unpacked the boxes. All the parts are available. Installed the machine. IQ, OQ and PQ performed and found satisfactory. Now the instrument is working fine. The Temperature and Vacuum is under range.					
NOTE:					
SPARE PARTS USED/ REQUIRED			TO BE FILLED BY CUSTOMER		
			Yes	INSTRUMENT INSTALLED AND FOUND WORKING SATISFACTORILY	
			NA	PREVENTION MAINTENANCE CARRIED OUT SATISFACTORILY	
			NA	FAULT RECTIFIED & INSTRUMENT WORKING SATISFACTORILY	
			NA	PLS AWAIT OUR WRITTEN APPROVAL FOR PARTS	
					COMMENTS (IF ANY)
ENGINEER'S SIGNATURE		NAME <i>Venkata Krishna</i> DATE <i>04-04-25</i>	 CUSTOMER'S SIGNATURE		
ONLY FOR PARTS CHANGED & ALL OTHER CHARGES WILL BE EXTRA AS APPLICABLE					
NOTE: PARTS REPLACED ARE CHARGEABLE EXCEPT DURING WARRANTY PERIOD. CONSUMABLES USED FOR THE INSTRUMENT ARE NOT COVERED BY WARRANTY & HENCE ARE CHARGEABLE. PARTS REPLACED DUE TO MISHANDELLING DURING OPERATION OR OTHERWISE WILL ALSO BE CHARGED IN EVERY CASE. IN CASE OF DIFFICULTY OR SUGGESTION PLEASE CONTACT SERVICE INCHARGE.					