



MICROFLOW DEVICES INDIA PRIVATE LIMITED

Regd. Office : Plot No : 4 & 5 WIP Sidco Industrial Estate, Thirumudivakkam,
Chennai - 600 044. Ph : 91-44-24781010 / 95000 82797
Correspondence Address : No. 52G, 1st Cross Street, Sabari Nagar, Porur,
Chennai - 600 116.



INSTALLATION / SERVICE REPORT

No. :

Date : 05/05/17


Client : Indian Immunologicals limited
Location : Rakshapuram, Kankapatta,
Contact Person : Mr. Nataraj
Department : _____


Nature of work : Installation Modification AMC
 Warranty General Service Others

| Model | Sr.No | *Service rendered | Average velocity fpm | Pressure mm of water |
|-------------|--------------|---|----------------------|----------------------|
| MFD/FT/1500 | 2585/02/2017 | Fume Hood Installation work is completed. | - | - |

Remarks

- 1) Sink provided by acrylic sheet but URI maintain epoxy
- 2) IQ, OQ, PQ & DQ not provided
- 3) ~~Activity~~ not properly sealed. Edges


9500120049
Rajesh P / B. Guesh Babu.
Customer Support Engineer

Client's Remarks

05-05-17
Date
Signature

* Above remarks completed on 16.06.17