

Service Report



BIODIGITAL (P) LTD.

Corporate Office

E-34, The Belaire, Golf Course Road, DLF Phase 5,

Sector- 53, Gurgaon 122002, NCR, Haryana

Tel: 0124-4065283 Fax: 0124-4147647

Service Type
<input checked="" type="checkbox"/> Installation
<input type="checkbox"/> Warranty Service
<input type="checkbox"/> Non Warranty Service
<input type="checkbox"/> AMC

Customer Details:

Murali Krishna

Name:	Mithun Kumar/	Date Reported:	07.09.2020
Institute:	IPL, Karapatala		
Address:	IPL, Karapatala		
Phone Number:	9100936576		
E-mail address:	Murali.krishna.b@indimmune.com		

Device Description

Instrument:	ged doc	Serial Number:	03000036/04000042
Model:	VUC/VUF	Date of Visit:	07.09.2020

Problem Found

Installation of ged doc

Service Performed

Ged doc has been installed. IR, OR, DR, PR has been performed based on the attached documents. Demo of the machine has been given as well. System is working absolutely fine. Images are good quality.			
Performed By:	MChresh Shetty	Date:	07.09.2020

Parts Installed/Removed

Part Name	Part Number	Serial Number	Replaced By
 			

Action Required

Instrument installed and demo performed. software installed printer installed. staining imparted

sholup
07.09.2020

Customer's Signature

07.09.2020
Representative