

SUN TEKNOLOZY / SUN LABORATORY SERVICES / SUN SCIENTIFIC CORPN

14/1 Lakshmi Dwarka Villa, Thalayari Street, West Mambalam, Chennai-600 0033, Ph.:044-24740143, Fax : 044-42024721
 Email: admin@sunteknozozy.com, sunteknozozy@gmail.com

SERVICE REPORT CUSTOMER DETAILS

05.10.17

Date: 06.10.17

Customer Name & Address	Contact Person	Mr. John	
HUMAN BIOTechnICALS	Designation	Executive	
INSTITUTE,	Dept.:	Micro Biology	
Kozhupannai, Pudumund (Po),	Tel. No.:	O.:	Extn.:
Udhagamandalam - 643 007	Mobile.:	8754574544	
	E-Mail.:	wjohn@indimmune.com	

INSTRUMENTS DETAILS

Instrument Name	Surface Air Sampler
Instrument Model	SAS Super ISO 100
Sr. No	16-03-3368

SERVICE DETAILS

Installation
 warranty
 AMC
 Services
 Calibration
 Chargeable/Free. if chargeable Amount Rs. - NA -

Problem Reported:
 New installation of Microbial Air Sampler ISO 100

Action Taken:
 Supplied the SAS Super ISO 100, Completed the installation of room air sampler. IQ, OQ and PQ document also completed. Given demo and training on instrument operation. Instrument is working in good condition.

SPARE PARTS REPLACED / RECOMMENDED

Sr. No.	Item Code	Item	Qty.	Replaced	Recommended
NA					
M. Suresh					

Customer's Remark: Qualification of SAS ISO 100 room air (microbial) sampler completed and working in good condition.

M. Suresh 06.10.17 Engineer's Signature	M. SURESH Machine Working Satisfactory C. JOHN WESLY KIRUBAKARAN 06.10.17 Customer's Seal & Signature
---	---