



### INSTALLATION REPORT

Installation Date	10-Jan-25	Spinco ID	S0001/33555	Job Status (Completed/Pending)	Completed
Instrument Model	Prominence P-Series	P.O.Number	4511005100	P.O Date	12-11-2024
Principals	SHIMADZU	Invoice No	LLP-2408-2024	Invoice Date	04-12-2024
Customer ID	AS/HPLC/S88	Other Details :			
<b>Account Information</b>			<b>Contacts (Manager &amp; Users)</b>		
Account Name:	Indian Immunological Ltd.	Name	D Mahesh	Designation	General Manager (GM)
Address & City:	Indian Immunological Ltd. Rakshapuram, Gachibowli, Hyderabad 500032	Mobile	9010081155	E-mail	dmahesh@indimmune.com
Phone :	9010081155				
Mobile :	9010081155				
Email :	9010081155				
Website :	www.ill.com				
<b>Pre-Installation Checks</b>					
Line - Neutral (230 V ± 10 %, ± 1 V)	230	Temperature (18 °C to 25 °C)			25
Line - Earth (230 V ± 10 %, ± 1 V)	230	Humidity (40 % - 70 %)			50
Neutral - Earth (< 1 V)	0.2	A.C (Yes/No)			Yes
Instrument Chassis to Ground (< 100 Ω)	50	Dust Free (Yes/No)			Yes
<b>Other Observations, If any</b>					
<b>Installation &amp; Demonstration (Yes/No)</b>					
1. Installation Stacking	Yes	1. Document Fill-up		Qualification (Yes/No)	
2. Flowline/Gasline Setup	Yes	2. Document Verification			Yes
3. Leak/Pressure Check	Yes	3. Is Verified Document Submitted			Yes
4. Cables Organization & Tying	Yes				Yes
5. Solvent / Drain bottles - Capped	Yes	<b>Other Details</b>			
6. PC Settings	Yes				
7. PC Check	Yes				
8. Software Check	Yes				
9. Hardware Check	Yes				
10. Demonstration	Yes				
<b>Date &amp; Time</b>			<b>Instrument Configuration</b>		
Date	Time In	Time Out	Model Name	Version	Serial Number
06-01-2025	10:30 AM	5:30 PM	LabSolution	6.124	L54746002934
07-01-2025	10:30 AM	5:30 PM	labSolution LC	L54746002934	IVJ
08-01-2025	9:30 AM	6:00 PM	PDA	L54746201620	JVZ
09-01-2025	10:30 AM	6:00 PM	DGU-2A5r		L22736257130
10-01-2025	10:30 AM	4:00 PM	CBM-20Alit	3.82	L20226291869
			LC-20AD	2.12	L20106298783
			SPD-20A	2.08	L22276210601
			SIL-20AC HT	2.15	L20356210642
			CTO-20AC	2.1	L20216240371
Engineer Observations :			Customer Comments :		
			<i>Installation and Demo completed successfully</i>		
Engineer Signature:			Customer Signature with Date & Stamp :		
<i>Pradeep G 10/01/2025</i>			<i>[Signature] 23-01-25</i>		
Engineer Name	Mr. Pradeep.G	Customer Name	D Mahesh		
Mobile	8925901877	Designation	General Manager (GM)		
E-mail	Pradeep.g@spincotech.com	Mobile	9010081155	E-Mail	dmahesh@indimmune.com