

SERVICE REPORT



CARE BIOSYSTEMS INDIA PVT. LTD.

8 & 8A, Ground Floor, Swaraj Daffodils, Plot No.96,
Sector 8A, Airoli, Navi Mumbai - 400 708. Maharashtra
State, India, Tel: Tel: +91-8655069660, +91-8655069661
Email : service.mumbai@carebio.com

Date : 17/06/2020 JC# 307

Engineer's name Anthony Reddy

Date of Request _____ Date of Visit 10,15,17/06/2020

P.O. No. : 4511003102, 12.07.2020

Inv. / Qtn. No. : _____

Customer's Name & Address

Indian Immunologicals Limited
Survey No. 281-284 & 321, Biotech Park Phase 2
Kavkapita, Telangana.

Type of Services

1 Installation	<input checked="" type="checkbox"/>	6. Paid Service	<input type="checkbox"/>
2 Breakdown	<input type="checkbox"/>	7. Calibration	<input type="checkbox"/>
3 AMC	<input type="checkbox"/>	8. IQ-OQ-PQ / Mapping	<input checked="" type="checkbox"/>
4. Warranty	<input type="checkbox"/>	9. Application	<input type="checkbox"/>
5. Inspection	<input type="checkbox"/>	10. Goodwill	<input type="checkbox"/>

Tel. No. : _____ Fax No. : _____

E-mail : _____ Estimated Amount Rs. _____ Add GST as applicable

Instrument	Model No.	Sr. No.
<u>PHCbi(Panasonic)</u> <u>Pharmaceutical Refrigerator</u>	<u>MPR-S313-PE</u>	<u>19070427</u>

Request for / Fault : Installation & Qualification

Supply Voltage : 228 V.AC Room Temp. : 21°C AC. : Yes No

Stabilized supply : Yes No Location : Room No. 02 Floor 1st

Instrument Status : Clean Average Not Good Needs Attn.

Set Temp. = 05°C Actual Temp. = 05°C Set RPM = - Actual RPM = -

Set = - Actual = - Set = - Actual = -

Work done : As per customer request visited location and completed the installation and qualification (2 hours without load and under load mapping). Refrigerator working satisfactorily. Submitted the qualification documents

Spare Parts Used / Required : _____

Notes / Conclusion : Supplied Datalogger, Printer, & Sensor.

Job Completed : Yes No

Customer's Remarks : Vendor validation completed.

Engineer's Signature Anthony Reddy 17/06/2020 Customer's Name & Signature : Venkat 17.06.2020