

FIELD SERVICE REPORT

eppendorf

Engineer's Name: B. SRIKANTH	S.S.R. No.: —	Date: —
Branch: HYDERABAD Mobile No.: 9121605485	W.o. No.: —	Date: —
E-mail: srikanth.b@eppendorf.co.in	Report Date: 05-04-2023	

No. of Day(s) worked: 01 From: 05-04-23 To: 05-04-23

Customer Address: Indian Immunologicals Ltd, R&D Lab, Rakshapuram, Gachibowli Hyderabad	Instrument(s) Details:		
	Model	Serial No.	Version
	Cryocube	F440LJ400800	
	F440n ID: R&D/VVIR/ULF/026		

Attention: Mv. Praveen. Designation: _____
 Phone: 9030465523 Fax: _____ E-mail: _____

MATERIAL USED		
Part No.	Description	Quantity
/	/	/
/	/	/
/	/	/

Problem Reported: **IQ/OQ ACTIVITY**

Details of work done:
 During visit unpacked the instrument and positioned at required location. 3pin top was changed, stabilizers is connected and power supply checked. Functional checks as per protocol completed and found Ok. Instrument is working fine.

Job Completed
 Job Pending
 Billing: O/W
 Warranty
 AMC
 Chargable

Action to be taken: — Recommendation: —

Customer Remarks:

Engineer's Signature <i>sn</i>	Customer's Signature with seal <i>Praveen</i> 03/05/2023
-----------------------------------	--