



## SERVICE REPORT



**CARE BIOSYSTEMS INDIA PVT. LTD.**

8 & 8A, Ground Floor, Swaraj Daffodils, Plot No.96,  
Sector 8A, Airoli, Navi Mumbai - 400 708. Maharashtra  
State, India, Tel: Tel : +91-8655069660, +91-8655069661  
Email : service.mumbai@carebio.com

Date : 21/08/2023

JC No.

346

Engineer's name

*S. Suresh Kumar*

Date of Request

Date of Visit 27/07/2023

MUM  AMD  HYD  PUN

Inv. / Qtn. No. :

Customer's Name & Address *G. Mohi Paf*

P.O. No. : 4513002852 04/01/2023

*Indian Immunologicals Limited  
Gachibowli, HYD.*

Type of Services

- |                |                                     |                       |                                     |
|----------------|-------------------------------------|-----------------------|-------------------------------------|
| 1 Installation | <input checked="" type="checkbox"/> | 6. Paid Service       | <input type="checkbox"/>            |
| 2 Breakdown    | <input type="checkbox"/>            | 7. Calibration        | <input type="checkbox"/>            |
| 3 AMC          | <input type="checkbox"/>            | 8. IQ-OQ-PQ / Mapping | <input checked="" type="checkbox"/> |
| 4. Warranty    | <input type="checkbox"/>            | 9. Application        | <input type="checkbox"/>            |
| 5. Inspection  | <input type="checkbox"/>            | 10. Goodwill          | <input type="checkbox"/>            |

Tel. No. :

Mobile No. : 9000225540

E-mail :

Instruments

Model Nos.

Sr. Nos.

*PHC6?  
UCT freezer*

*MDF-U55V-PE*

*22050026*

*ID NO:- A5/ULF/578.*

Request for / Fault :

*Installation, IQ, OQ, PA mapping*

Mains Supply : 231. V.AC Current : 3.8 Amps Enviro.Temp : 18.0 °C Aircon : Yes  No

Stabilized supply : Yes  No

Location : Room No. *Incubator & freezer room* Floor *2nd*

Instrument Status : Clean  Average  Not Good  Needs Attn.

Set Temp. = *-70.0°C* Actual Temp. = *-70.0°C* Set RPM = *-* Actual RPM = *-*

Set = *-* Actual = *-* Set = *-* Actual = *-*

Work done : *Installed the freezer. Achieved set temperature performed IQ, OQ, PA as per customer URS. Demoted the customer. Submitted all documents. freezer is running good Calibration.*

Spare Parts Used / Required :

Notes / Conclusion :

*Freezer is working Satisfactorily.*

Customer's Remarks :

Job Completed :

Yes  No

*Sig. 21/08/2023*  
Engineer's Signature

*Mohi Paf*  
Customer's Name & Signature