



**Spectralytic Scientific India Pvt. Ltd.**

Unit no.005B 8th Floor, G corp Tech Park,Ghodbunder Road  
Kasarvadavli,Thane(west) - 400607. INDIA.  
Tel.: 022-33261700

**FIELD SERVICE REPORT**

CIN:U74999MH2004PTC145744

<b>CUSTOMER NAME</b>	Indian Immunologicals Limited	<b>ENGINEER</b>	: V. Abhilash
<b>ADDRESS</b>	Karkapatla	<b>DATE</b>	: 28-Jul-23
	Turkapally, Hyderabad	<b>PROD. LINE</b>	: Mat Char, FTIR
<b>CONTACT PERSON</b>	MR. K. KISHAN	<b>MODEL NO.</b>	: Spectrum Two
<b>DESIG. &amp; DEPT.</b>	QC Head	<b>SERIAL NO.</b>	: 92204
<b>MOBILE</b>	8143260036	<b>JOB CODE</b>	: NA
<b>EMAIL</b>	k.kishan@indimmune.com		

**FAULT REPORTED :** FTIR Instrument Software Upgradation



Checked and Verified the Computer Configurations, found ok  
then performed and Verified Software Installation, found Software installed correctly without any error, then performed the Software  
Operational Qualification, found all the test parameters are passed, filled OQ documents and submitted to end user  
then performed Instrument OQ, found test paramters are passed, then given Software operating procedure demonstration to end users & QA  
clarified there Quarries, created report templates as per there requirement.  
Found Software & Instrument Working in good condition.  
Given Demonstration to end users.

**JOB COMPLETED**

YES  NO

Remarks:

NO.	PART DESCRIPTION	PART NO.	REPL PEIL	QTY.
NA	NA	NA	NA	NA
NA	NA	NA	NA	NA
	NA			
	NA			

DATE	TRAVEL TIME	TIME ON JOB		For Spectralytic Scientific (I) Pvt.Ltd.	FOR CUSTOMER
		START	END		
26-Jul-23	01hrs	11:30	17:30	 Digitally signed by V ABHILASH Date: 2023.07.28 17:03:32 +05'30'	
28-Jul-23	01hrs	11:00	17:00		
	NA			28-Jul-2023	28-Jul-2023
	NA			SIGNATURE	SIGNATURE