

# INSTALLATION REPORT

inkarp®

H.O.:1-2-45/1, Street No.2, Kakateeya Nagar Colony, Habsiguda, Hyderabad-500 007. T.S. India. Tel.: 040-27172293 / 5088  
Fax: 040-27154686 E-mail: sales@inkarp.co.in Website: www.inkarp.co.in CIN No.: U33112TG2004PTC043966

Report No. \_\_\_\_\_ Date : 07/08/2020

Name	Mr. Sathesh.
Institute / Industry	IIL
Department	Instrumentation Lab
Place	Gachibowli

PO No. 4513002438 Our Invoice No.:

Principal	Instrument Description	Serial No.
HITACHI	NEXTR STA 200	2011205807-02


Warranty Valid Upto : one year Agency Commission Applicable : Yes / No


Job Done

- consignment received with out any damage.
- All Accessories received as per Po.
- Instrument Install the correct place and install the software.
- checked the communication and found ok.
- we performed baseline calibration and standards Indium, Tin calibration.
- Demo and sample run in pending.

Customer Remark

Demo and sample run has to be performed yet

  
07/08/2020  
Name & Signature of Customer  
with seal

  
07/08/2020  
Name & Signature of Engineer  
M. Sai Reddy  
N. Vivek Reddy  
(331143281)

\*\*Please use additional sheets if required.