

Sartorius Stedim India Pvt. Ltd.
 #69/2 & 69/3, Kunigal Road,
 Jakkasandra, Nelamangala,
 Bangalore 562123
 Phone: +9180 .68154177
 E-Mail: CustomerService.India@Sartorius-Stedim.com
Service Report: #2801



CUSTOMER
Customer No.: 20013554 Indian Immunologicals Limited Rakshapuram, Gachibowli Post, Hyderabad 500032 India

Additional Information
PO No.: 4511003535 Service Contract: None Complaint No.: Type: Installation Technician: Kunchala Srihari, Ramakanth Seelam


SERVICE NOTES
Installation for Sartocheck 5 Filter Tester Sartocheck 5 Filter Integrity(41507500) Installation and Qualification activity is completed. Operational demo and training completed. Unit and all accessories(YDP printer, paper role, stabuli, power cables, LAN cable, screw driver, Inlet and outlet tubes, USB cable, Barcode scanner) handed to customer. Unit is working good . Unit and Qualification (IOQ) documents handed to customer. Customer Details: Name: P.Sukumar Contact no:9490065900 Mail ID: psukumar@indimmune.com


EQUIPMENT				
NAME	LOCATION	MFR.	MODEL NO.	SERIAL NO.
Sartocheck 5 Filter Tester	20013554		26787--FT	41507500

Labor				
Date	Item No.	Description	QTY	Comment
Jul 5, 2021		Travel Time	3.00	Travel time
Jul 5, 2021		Working Time	8.00	Working hours
Jul 6, 2021		Travel Time	1.00	
Jul 6, 2021		Working Time	7.00	
Jul 12, 2021		Travel Time	1.00	
Jul 12, 2021		Working Time	1.25	

Material				
Item No.	Description	QTY	Comment	

Expense Total				
Date	Item No.	Description	QTY	Comment

CUSTOMER NAME: (P Sukumar)
Customer Signature 
DATE Jul 12, 2021 @ 3:46pm

TECHNICIAN NAME: (Ramakanth Seelam)
Technician Signature 
DATE Jul 12, 2021 @ 3:46pm