

PARTS (CONTINUED)

ORDERING PARTS AND CONSUMABLES

If you have the part number for an item, you may order it directly from Sheldon Manufacturing by calling 1-800-322-4897 extension 3. If you are uncertain that you have the correct Part Number, or if you need that specific item, please contact Sheldon Technical Support for help at 1-800-322-4897 extension 4 or (503) 640-3000. Please have the **model number** and **serial number** of the BACTRON ready, as Tech Support will need this information to match your workstation with its correct part.



SERVICE REPORT

(PHONE VISIT)

From (Name): Shyam Vrasod / Place: Hyderabad Date: 05/08/16

Installation / AMC Service / Repair (Tick whichever is Applicable)

CUSTOMER (Organization + Division/Dept. + Place) : Indian Immunologicals Limited
Dept. of R2D, Roli shapuram,
Gachibowli Post, Hyderabad - 500032

In-charge of Equipment : Dr. M. Elaiyaran Phone No. : 09010540542
 e-mail ID : _____

Customer's Purchase Order No. / Complaint : _____

Make + Equipment / Name + Model : Shellab + Anaerobic chamber + Bactron E2-2

Principals Invoice No. / Machine Serial No. / Year Of Supply : _____ / 103023016 / 2016

Type of Control Unit / Version (If any) : _____

NOTE ! NOTE ! : CHECK & ENSURE correct power supplies Before attending Instruments !!!

REPORT : 230V 80V

1) DATE of Complaint / Confirmation of availability of Installation Pre-Requisites : _____ / _____ / _____

Message received from : Mr. / Ms. Dr. Elaiyaran

2) PROBLEMS / COMPLAINTS Received / Observed ; if any :

* Installation.

3) WORK ATTENDED (Description ; Including for Installation) :

- checked supply voltage
- Fitted all assy and connected Air gas cylinder.
- Run "establish anaerobic atmosphere" cycle of the chamber and established anaerobic inside the chamber.
- Given training on operation of the instrument and using of accessories.
- installation completed and handed over the instrument to customer.

Hours Spent : _____ Hrs. From : _____ Hrs. Date 05/08/16 to _____ Hrs Date 05/08/16

4) Has the Installation / Service Job / Repair Job been Completed ? Yes : No :

5) If the Work is discontinued, Why ? :

6) Your Suggestions :

7) If the customer has any comments to make: _____

8) Signature (Service Engineer)

Shy

9) Customer's Signature

Shyam Vrasod
05/08/16
2016

Name: A. Phanikumar Designation _____ Phone No. _____
 Sign of Scientist / Professor / Engineer (In-charge of Equipment)