

Installation Report

Customer Name:	Indian Immunologicals Ltd	Installation Date:	07/03/2019
Site/Dept:	Human Biologicals Institute	Warranty Start Date:	07/03/2019
Address:	Kozhipannai, Pudumund P.O	Warranty End Date:	06/03/2020
City:	Udhagamandalam	Report No.:	G7-3829495
Pin code:	643007	Engineer:	Lakshmi Narayanan / Anantha Raj
Contact:	Dr Cristoper / T. Shanmugasundaram	Mobile:	9840127339 / 9840127339
Designation:	Scientist	Email:	narayanan.k@eppendorf.co.in
Email:	-		
Phone:			
Product:	5424R Micro Centrifuge - Foil Version	PO No.:	4513002226
Serial No:	5404HN532009	PO Date:	11/01/2019
Software Version:	V2.9		

Details of Work Done


<input checked="" type="checkbox"/> Package is received in good condition <input checked="" type="checkbox"/> All the items were received as per PO <input checked="" type="checkbox"/> Customer power supply is OK <input checked="" type="checkbox"/> Instrument is installed at the suitable bench <input checked="" type="checkbox"/> Instrument operation is demonstrated <input checked="" type="checkbox"/> Instrument maintenance is demonstrated	<input checked="" type="checkbox"/> Instrument is working satisfactorily <input type="checkbox"/> epPoints registration done Username of epPoints account: <input type="checkbox"/> Product Registration Completed <input checked="" type="checkbox"/> Regulatory certificate details are explained
--	---



Scientist Trained:

Calibration Services according to ISO/IEC 17025:2005

Calibration parameters

Volume: 10µL to 10 ml |
 Speed: 25 rpm to 30,000 rpm |
 Temperature: -195 ° C to +200°C |
 Time: 60 seconds to 3600 seconds
 Instruments: Centrifuges | Incubators | Incubator shaker
 | Freezers | Pipettes | Dry/Water bath


CERT. No.
C-1159 / 0773 / 1160

Customer Remarks:	Engineer Remarks: Installation completed Instrument working satisfactory
Customer Seal with Signature:  07.03.19	Engineer Signature:  7/3/2019

