

SERVICE REPORT

NAME : Indian Immunologicals Limited
Rakshapuram, Gachibowli
Hyderabad

S.No. 25552 DATE 30-10-2015

ENGINEER : Vara Prasad TSV

MODEL NO / CAT NO : JLS-5000T

SERIAL NO : 15071648(P), 15071881(DC)

OTHER DETAILS OF THE UNIT : 15071717(Ac-AP)

PERSON : Dr. Rajanathan

INSTRUMENT STATUS: Warranty Contract Billable

/ FAX :

CALL REASON : Repair Courtesy Validation Installation

REASON : Installation and operational qualification


- SERVICES RENDERED :
1. Installed the Chromeleon software and instrument.
 2. Done the qualification of Chromeleon software and established communication with software.
 3. Done the qualification of Conductivity, ElectroChemical and VMD detector.
 4. Done the documentation of the qualification.
 5. Installation and operational qualification of the instrument completed successfully.


PART#	SPARES DESCRIPTION	QTY. REPLACED	REPLACED FROM (✓)		QTY. REQUIRED
			CUST. STOCK	THERMO STOCK	
<u>NA</u>					

REMARKS : IQ/OQ Completed.

JOB COMPLETED : YES NO

From	To	Hrs. Worked


(Chozhavu Rajanathan) 31.10.15
CUSTOMER'S FULL NAME/SIGNATURE STAMP


Vara Prasad TSV 31-10-2015
ENGINEER'S/SIGNATURE FOR THERMO - INDIA

ACTION REQUIRED: YES NO

a) No. _____ b) SAP Notification No. _____ c) SAP Equipment No. _____