

"Metrohm-SIRI Towers", 3&4 Floor, Fourrts Avenue, Annai Indira Nagar, Okkiyam, Thoraipakkam, Chennai – 600097.

Tel: +91 44 40440440 Email: info@metrohm.in Web: www.metrohm.in

Engineer's Name:

Activity Date(s):

Report Date:

Sathes Kumar.K

29/05/2023 & 30/05/2023

30/05/2023

Customer Address :

Instrument(s) Details

Human Biological Institute
(Indian Immunologicals limited)
Ooty

Principal

Model

Serial No.

Software Ver

Metrohm

890 Titzevole

20447

tiam030

Metrohm

860 KI Thermo

52305

S.860-0011

Prep

Attention:

Designation:

E-mail:

Contact Details :

P.O No.: 4511004259

dtd. 02/12/2022

Invoice No. : 141-1032730

dtd. 03/03/2023

Details of Work done:

checked the consignment against the packing list, found all the parts are received in good condition. Installed the instrument along with necessary accessories. Done IO,OP for the instrument. All the IO,OP documents are signed and given to the end user. Instrument working in good condition.

Details of Application:

Moisture Content determination in sample

Person (s) Trained:

Mr. Balaramanugan, Mr. Ramesh and Mr. Neelhan.

Action to be Taken / Engineer's Remark :

Demo and training about instrument operation and maintenance has been given to the end user. Need to change the 2ml vial adaptor and customer sample testing is pending.

 Job Completed Job Pending

Customer Remarks: demonstration done & instrument installed. Warranty certificate yet to be provided.

Visit Date

Start Time

End Time

Rating :

Poor

Fair

Good

Very Good

Excellent

 1 2 3 4 5 6 7 8 9 10

Engineer's Signature



Customer's Signature with Seal


