

Service Report

Date : 23 May 2022



Company :	INDIAN IMMUNOLOGICALS LIMITED GACHIBOWLI POST RAKSHAPURAM HYDERABAD 500032 INDIA	Work Order No.:	7000567354
Contact Name:	Rama Krishna	Purchase Order No.:	NA
Contact Phone:		Service Address:	Survey No 281-284&321 Biotech Park, Phase-III Karkapatla vill SIDDIPET 502281 INDIA
Contact Email id:	p.ramakrishna@indimmune.com		

WORK ORDER DETAILS		
Service Type	Serial Number	Registered Product Description
Installation / IQOQ	14833044	Flowstar IV Standard

Items

Item/Po Item	Product	Description	Quantity	Field Service Engineer
10	PTS02QSFSTAR	Flowstar IQ/OQ at Site	1	Mohan Kurukundu
20	ISGFLAB	FIELD LABOUR (HOURS)	6	Mohan Kurukundu

Symptom Reported:
None
Problem Observed:
None
Service Performed:
1.Completed the IQ/OQ of the instrument as per Pall document no GGFD0422 and all tests are passed successfully.2. Instrument and External vent valves are found working fine. 3.Hand overed the filled documents to customer.
Verification:
Date of Visit: 22nd April 2022

	23 May 2022		23 May 2022
Customer Signature:	Date	Engineer Signature: Mohan Kurukundu	Date