

SERVICE REPORT

8247861566



CARE BIOSYSTEMS India Pvt. Ltd.

8 & 8A, Ground Floor, Swaraj Daffodils, Plot No.96,
Sector 8A, Airoli, Navi Mumbai - 400 708. Maharashtra
State, India, Tel: Tel : +91-8655069660, +91-8655069661
Email : service.mumbai@carebio.com

Date : 12/02/2020 JC# 221

Engineer's name *Anthony Reddy*

Date of Request Date of Visit 31/01/2020
04.03.12/02/20

P.O. No. : 4511003012 , 24/12/2019

Inv. / Qtn. No. :

Customer's Name & Address
*Indian Immunologicals Limited.
Rakshapuram, Gachibowli
Hyderabad. - 500032.*

Type of Services	
1 Installation <input checked="" type="checkbox"/>	6. Paid Service <input type="checkbox"/>
2 Breakdown <input type="checkbox"/>	7. Calibration <input type="checkbox"/>
3 AMC <input type="checkbox"/>	8. IQ-OQ-PQ / Mapping <input checked="" type="checkbox"/>
4. Warranty <input type="checkbox"/>	9. Application <input type="checkbox"/>
5. Inspection <input type="checkbox"/>	10. Goodwill <input type="checkbox"/>

Tel. No. : Fax No. :

E-mail : Estimated Amount Rs. Add GST as applicable

Instrument	Model No.	Sr. No.
<i>Panasonic ULT Freezer</i>	<i>MDF-U55V-PE</i>	<i>17070200</i>

Request for / Fault : *Installation & Qualification.*

Supply Voltage : *229* V.AC Room Temp. : *20* AC. : Yes No

Stabilized supply : Yes No Location : Room No. *R&D* Floor *Ground.*

Instrument Status : Clean Average Not Good Needs Attn.

Set Temp. = *-70°C* Actual Temp. = *-70°C* Set RPM = Actual RPM =

Set = Actual = Set = Actual =

Work done : *Asper customer request visited site and installed the ULT Freezer and Completed the IQOQ&PQ (72 hours without load & with load Mapping). freezer working satisfactorily*

Submitted the Qualification Document.

Spare Parts Used / Required : *Nil*

Notes / Conclusion :

Job Completed : Yes No

Customer's Remarks :

Engineer's Signature *Anthony Reddy
12/02/2020*

Customer's Name & Signature : *Chakraborty
12.02.2020*