

Microflow Devices India Private Ltd.

No. 51G, 1st Cross Street, Sabari Nagar, Porur, Chennai - 600 116 India. Phone : +91-44-4284 0251

SERVICE REPORT

Client : Indian Immunologicals htd
 Location / Dept. : Kachibowli, Hyderabad
 Person contacted : Mr. Phanikumar
 Department : R&D, (media preparation lab)

No. :
 Date : 08/08/13

Nature of Work : Installation Modification Others
 Warranty Special

Model	Sr.No.	*Service rendered	Average Velocity fpm	Pressure mm of water
MFD-V-1800	1279-12-2012	6" Vertical laminar Air flow installation and testing was completed. (PAO, velocity and Particle Count test)	F1 - 99 F2 - 102	12

Remarks: All the parameters are within the limits. laminar flow working good condition.

*A : Prefilter B : Manometer
 C : Motor - blower D : Lights

Client's Remarks

Prakashan,
 Customer Support Engineer

Prakashan,
 Date: 08/08/13
 Signature