



Microflow Devices India Private Ltd.

No. 51G, 1st Cross Street, Sabari Nagar, Porur, Chennai - 600 116 India
 Phone : +91-44-4284 0251



SERVICE REPORT

Client : Human Biologicals Institute No. : _____
 Location / Dept. : Cachibowli, Hyderabad Date : 18/04/14
 Person contacted : Mr. Panikumar
 Department : R & D Dept
 Nature of Work : Installation Modification Others
 Warranty Special

Model	Sr.No.	*Service rendered	Average Velocity fpm	Pressure mm of water
MFD-SS VLAF- Y-1880	2191 -12- 2014	vertical laminar air flow installation has been completed.	-	-

Remarks All the parameters are within the limits, laminar flow working good condition.
 * validation to be done.

* A : Prefilter C : Motor - blower	B : Manometer D : Lights	Client's Remarks
Arshman. 		
Customer Support Engineer	Date 18/04/14	Signature