



SERVICE REPORT

mk technology1 Pvt. Ltd.

HO : GG-1, 139-C, PVR Road, Vikas Puri, New Delhi-110018 INDIA

☎ +91 9810-55-0075 / +91 9958-20-9029 ✉ info@teknology1.com

Report No. **2105**

Report Date: **03/01/22**

Customer details	Instrument details		Work details	Order details
Company: Human Biological Inst (IIL)	Make / Model	CLIMET CI-90A	Installation <input checked="" type="checkbox"/>	New Sales <input checked="" type="checkbox"/>
Address: Coaty, TN-643007	Serial No.	216500 216501	Calibration <input type="checkbox"/>	Warranty service <input type="checkbox"/>
	Calibration date	22/04/2021	Up gradation <input type="checkbox"/>	Work order <input type="checkbox"/>
	Calibration due date	22/04/2022	Break down <input type="checkbox"/>	AMC <input type="checkbox"/>
	Instrument ID		Validation <input type="checkbox"/>	Gratis <input type="checkbox"/>
	Sensor details	MAS	PM <input type="checkbox"/>	Others <input type="checkbox"/>
	Software/ version		Training <input type="checkbox"/>	Job Performed at
	Firmware version		Inspection <input type="checkbox"/>	Customer site <input checked="" type="checkbox"/>
			Pre Installation <input type="checkbox"/>	mk technology1 <input type="checkbox"/>
			Others <input type="checkbox"/>	Order No:

Objective: **Installation of 02 units of CI-90A MAS.**

Service performed :

- 1) Boxes unpacked
- 2) Physical run of 02 units of CLIMET CI-90A
- 3) Setup of 02 units of MPD for use in Comp. Air.
- 4) Settings are done as per user requirements.
- 5) IQ / OQ done

Customer's remarks: **As Committed Equipment Calibration due April 2022 shall be free of cost.**

Engineer's remarks: **Installation done & working perfectly**

Proof of work

Spare parts details

Date	Time in	Time out	S. No.	Description	Part No.
27/12/21	01:00pm	5:30pm			

Engineer: **Supriyo Roy** Instrument is handed over in satisfactory condition.
 Date: **27/12/2021** User name: **Mr. Satish**
 Sign: **[Signature]** Sign/ Date: **[Signature] 03/01/22**
 Email ID: **P.Satish@indimms.com**
 Contact No: **9942538587**

Particle Measurement, Clean Room Monitoring, Filter Testing, Clean Utilities Testing and PAT Instruments
 Compressed Air/Nitrogen & Pure Steam Validation

D. Vengatesan (8012755566) 2nd User