

MOM ON CIRCULAR CHART RECORDERS

Date: 17-04-2017

- **Vendor:** SunSui Process Systems
- **Ref.Po.No.:** 4511002233
- **Location/Plant:** HBI OOTY
- **Machine Specification:** Circular Chart recorder
- **Model:** SCR-60
- **Machine Sr.no:** 17040001

Work done Details:

1. One no of the recorder bearing the serial no 17040001 installed as advised by the User at in -70 c Ultra freezer
2. The remaining Five (5) recorders handed over to HBI stores.
3. One of the Engineering person was trained on the programing sequences


Follow ups/Pending activity:


1. If required necessary support should be provided for the installation of the remaining five recorder.
2. The vendor shall supply 25nos of Daily charts for all the instruments.
3. The Warranty period is calculated from the date of supply of the instruments at HBI Ooty.
4. Spares and accessories lists are to be provided for the instrument.


Conclusion :

The installed circular chart recorder will be under observation for a period of one month, on successful completion of the due period the installation and commissioning is termed as completed


For Human Biologicals Institute


(R.Suresh-HBI engineering)


(R.H.Kingsley-HBI engineering)


(C.Ganesh-HBI engineering)

For SUNSUI PROCESS SYSTEMS


(Rakesh Bhuvad-Engineer)

9637802111