



Spinco Towers, Plot No-2,
Block-D, Street No-3, Kakatiya Nagar,
Habsiguda, Hyderabad-500 007.
Ph No: 040-27175520/21/22

INSTALLATION REPORT

CRM PM No :		Spinlab ID: B007/35	Engineer Name: P Satyasagar	
Account :	Human Biological Institute (Division of IIL)	Customer ID:	P.O.No. :4513001920	Date: 23 feb 2017
Address :	Kuzhipannal, Pudumund Post, Udthagamandalam, Ooty	Principal: BW Tek Inc	Invoice No: #URS/QC/HRS/011	
State :	Tamilnadu	Pin Code:	643007	Job Completed : Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>
SID In-Charge Name :	Dr. D. Vengatesan	System User Details		
Designation :	Deputy Manager	Name	Designation	Mobile
Department :	Quality Control	Mr. Ramesh	Executive	9944612976
Phone :				E-Mail
Fax :				v.ramesh@indimmune.com
mobile :	8012755566			
Email :	vengatesan@indimmune.com			
Website :	www.indimmune.com			

Activity
Installation of NanoRam Instrument (SANXOI)

- Actions Taken**
1. Checked the consignment and found no short shipment.
 2. Installed the instrument and the software.
 3. Performed the IQOQPQ of the NanoRam Instrument as per protocol, all tests are passed without any deviations.
 4. Trained the users on how to use the instrument, Software and how to create a method by creating method of Human Serum Albumin.
 5. Handedover the document to the user and instrument is working fine.

AC Line Voltage Reading : P - N = 221 V P - E = 216 V E - N = 5 V UPS Stabiliser

Further Action Required :

Customer Comments :

Date	Time IN	Time OUT	Engineer Signature :
22-Feb-17	10:00	18:30:00	<i>[Signature]</i> 23.02.17
23-Feb-17	9:00	17:00	
			Customer Signature : (With Seal)
MODEL	Sr.NO	Version	<i>[Signature]</i> 25.02.17 (D. VENKATESAN)
BWS456-785	SANXOI		
NID Client Server		NIDV5.14 NR2	
			Date : 23-Feb-2017