

Service Report

Date : 23 May 2022

Company :	INDIAN IMMUNOLOGICALS LIMITED GACHIBOWLI POST RAKSHAPURAM HYDERABAD 500032 INDIA	Work Order No.:	7000567099
Contact Name:	Rama Krishna	Purchase Order No.:	NA
Contact Phone:		Service Address:	SURVEY NO. 281-284 & 321 MEDAK 502281 INDIA
Contact Email id:	p.ramakrishna@indimmune.com		

WORK ORDER DETAILS

Service Type	Serial Number	Registered Product Description
Installation / IQOQ	26836944	Flowstar IV Standard

Items

Item/Po Item	Product	Description	Quantity	Field Service Engineer
10	PTS02QSSTAR	Flowstar IQ/OQ at Site	1	Mohan Kurukundu
20	ISGFLAB	FIELD LABOUR (HOURS)	6	Mohan Kurukundu

Symptom Reported:

None

Problem Observed:


None

Service Performed:

1. Completed the IQ/OQ of the Instrument as per Pall Document no GGFD0422 and all tests are passed successfully. 2. Instrument and External vent valves are found working fine. 3. Hand overed the filled IQ/OQ documents to customer.

Verification:

Date of Visit: 22nd April 2022

	23 May 2022		23 May 2022
Customer Signature:	Date	Engineer Signature: Mohan Kurukundu	Date