

INSTALLATION / SERVICE REPORT



ULTRA INSTRUMENTS Plot No. 92, Amarjyothi Colony, Bowenpally, Secunderabad - 500 011. Tel: (9140) -27951055, 27951066 Email: us@ultragroup.in		Ref. No.: 106 Customer Service Order No.: _____ Installation <input checked="" type="checkbox"/> Breakdown <input type="checkbox"/> P.M.S. <input type="checkbox"/> General Visit <input type="checkbox"/>	
Customer : <u>IIL - H.B.I</u> <u>Block - B-5</u>		Inv No. _____ Dt. _____ PO.No. _____	
Contact : <u>MY.M.SURESH</u> <u>P. SOKUMAR</u>		Installation Date <u>08.10.2020</u>	
Phone/Mobile : _____		E-mail ID : <u>psukumar@indium.com</u>	
Product : <u>PH METER</u> Sl.No. : <u>19410003</u> Make : <u>OHAS</u> Car / Model <u>ST500</u>		Remarks : <u>Installation.</u> <u>Instrument handed over</u> <u>to customer in with</u> <u>Accessories in good condition</u> <u>with 225 V-Ac / 0.5A power</u>	
Diagnosis / Action Taken <u>Installation of OHAS PH</u> <u>meter is completed. Training also given to</u> <u>user. Calibration done with PH7 & PHA buffers</u>			
Completed <input type="checkbox"/>		Repeat Visit <input type="checkbox"/> <u>Test found okay</u>	
ENGINEER'S REMARKS		CUSTOMER'S REMARKS	
<u>Instrument working</u>			
<u>satisfactory.</u>			
Engineer's Signature <u>V. Satish</u> Engineer's Name <u>V. SATISH</u>		Customer's Signature <u>[Signature]</u> With Seal <u>08.10.20.</u>	