

INSTALLATION REPORT

End User's Name	Mr.P.Suresh Kumar		
Designation	Senior Executive	Deptt.	Production
Company	Indian Immunologicals Limited		
Address	Gachibowli ,		
	Rakshapuram	City	Hyderabad
	State	Telangana	Pin-code 500032
Contact No.		Mob.	9908886785
e-mail Id	psureshkumar@indiammune.com		

Date of Visit	22.03.2019	Reference	Mail Confirmation
Type of Call	Pre-Installation		Installation <input checked="" type="checkbox"/>

Instrument Details	
Type Of Product	Peristaltic Pump with Pump Head 77602-10
Catalog No.	77420-10 & 77602-10
Model	77420-10 & 77602-10
Sr. No	L18003394 & M18003507

Initial Check list		
Damage to Instrument	Yes	No <input checked="" type="checkbox"/>
Damage to parts	Yes	No <input checked="" type="checkbox"/>
Part Missing (Check against packing list)	Yes	No <input checked="" type="checkbox"/>



Description: Installation and Demonstration of the Peristaltic Pump

Any additional comments: Instrument along with all standard Accessories Received in good condition. Instrument installed, and Demonstration was given on the operation and Calibration. Performance Verified. Instrument is working fine.

Status Of Job	Completed <input checked="" type="checkbox"/>	Pending
---------------	---	---------

Engineer's Remark: Instrument is working fine.

Customer's Remark: we have observed that instrument damaging the silicon tube while running the pump at maximum flow rate.

Engineer's Name: Mr. Rajashekar.Potha	End-User's Name: Mr.P.Suresh Kumar
Signature:  22.03.2019	Signature:  22.03.19.

Cole-Parmer India Pvt. Ltd.

403, A -Wing, Delphi, Hiranandani Business Park, Powai, Mumbai 400 076
 Tel: 91-22-61394444 Fax: 91-22-61394422 # E-Mail: service@coleparmer.in