

Date: 10.12.2020

Report No. GTX/GIP/11L-001  
 (For Office Use Only)

### Performance Letter cum Installation Report


Order No.: 4511003344 dated 28.10.2020  
 User's Name: P. Suresh Kumar  
 Institute/Organization: Indian Immunologicals limited  
 Department: HH-production  
 E-Mail: psureshkumar@indimmune.com  
 Telephone No(s): 9908886785

Detail of Instrument(s) : Model & Make	Serial No.
peristaltic pump 50ml to 5000ml/min F6-6L & SHENCHEN (Genetix Brand)	00011010100057

Nature of Job: Warranty..... Installation ..... AMC.....

Job Description/ : The peristaltic pump demonstration & installation has been completed

User's Remark : The peristaltic pump installation completed. Provide tube clamp for 24 mm size.

Signature of User : 

For Genetix Biotech Asia (P) Ltd

Stamp and sign

Stamp and sign   

Print Name of user .....

Print name : VISWANATHAN

Office Remarks : Machine should be use with Online UPS