

MEDISPEC (I) LTD

Offering Product for future of Science ...

SERVICE REPORT

CL#	Date: 17.04.2019
-----	------------------

AHMEDABAD BANGALORE CHANDIGARH CHENNAI DELHI HYDERABAD KOLKATA LUCKNOW MUMBAI PUNE

Institution: Indian Immunologicals Limited, Karakpatla.	Attn: Mr. Arun & Mr. Sandeep Patel
	Department: HHQC
	Tel: +91 9553331100
	Email: j.arun@indimmune.com

Order Ref: PO. NO: 4513002248 dated: 04-02-2019

Model: Epoch 2NSC	Serial No: 1903116	Base code: 1770200	Principals <input checked="" type="checkbox"/> BioTek <input type="checkbox"/> Nickel <input type="checkbox"/> NuAire <input type="checkbox"/> UVP <input type="checkbox"/> GeSiM <input type="checkbox"/> Promicol <input type="checkbox"/> Biospherix <input type="checkbox"/>
Gen 5 V3.05.11	XH2M RWDB	Version: 1.09	
	CBBH XD	Assaycode:	
		Version:	

Service Complaint: Installation of the BioTek Epoch 2NSC microplate Reader.	Service Type <input type="checkbox"/> Inspection <input checked="" type="checkbox"/> Installation <input type="checkbox"/> U / Warranty <input type="checkbox"/> U / AMC <input type="checkbox"/> Preventive Maintenance <input type="checkbox"/> Paid / Service order <input type="checkbox"/> Add-on / Complementary
--	--

Observation: -	
-------------------	--

Service Details: consignment verification, Installation and calibration completed. IQ-OQ-PQ documents completed and submitted.	Utilities Check List <input checked="" type="checkbox"/> Electric Supply : 230VAC <input checked="" type="checkbox"/> Ground / Earthing <input type="checkbox"/> Voltage Stabiliser / UPS <input checked="" type="checkbox"/> Air Conditioning <input checked="" type="checkbox"/> Dust Free Environment <input type="checkbox"/> Computer <input type="checkbox"/> Printer <input type="checkbox"/> Interface : Serial / USB <input type="checkbox"/> COM Port : <input type="checkbox"/> Gas Supply <input type="checkbox"/> Distilled Water <input type="checkbox"/>
--	--

Part No.	Description	Qty	Invoice Y/N
/	/	/	/
/	/	/	/
/	/	/	/

Test Result: The instrument installed and working Satisfactorily.
--

Note / Remarks:

For MEDISPEC (I) LTD B. Murali 17.04.2019 Engineer's Name:	Customer's Approval / Acceptance: Signature & Seal: 17.04.19
---	--