

INSTALLATION REPORT

REPORT NO.



ABBOTT HEALTHCARE (P) LTD
DIAGNOSTICS DIVISION

Abbott Healthcare Pvt Ltd
Floor 13-19, Godrej BKC, Plot C-68
G-Block, Bandra Kurla Complex
Near MCA Club, Bandra East Mumbai -400051

Ph: 022-28717462, FAX: 022-28717499

Customer Name & Address :
INDIAN IMMUNOLOGICALS LTD
HYDERABAD

MODEL:
ARCHITECT I1000SR

SERIAL NUMBER: I1SR66538

WORK STARTED

WORK ENDED

INSTALLATION REQUIREMENT CHECK

DATE: 14/04/2025

DATE: 14/04/2025

UPS/REQUIRED VOLTAGE: 230 V

EARTHING VOLTAGE: 1 V

TIME:

TIME:

SPACE AVAILABILITY AS PER SPECS: YES NO

ROOM TEMP AS PER SPECS : YES NO

HARDWARE CALIBRATIONS / SETTINGS CARRIED OUT:
YES NO

ACCESSORIES RECEIVED

INSTRUMENT

SCC

MONITOR

ACC KIT

Assay Calibrated

HBsAgQ2

Architect instrument verified and installed in the lab

CUSTOMER REMARK

Instrument installed and working fine

COMMENT:

IB, OA, PA completed

CUSTOMER SIGN

: *V. Jalun*

F.S.E'S NAME: Mahesh Lokam

SIGN:

Mahesh

DATE

: *14-04-25*

DATE: 14/04/2025